



Title	Multiwalled carbon nanotubes coating accelerates osteoconductivity of anodized titanium [an abstract of entire text]
Author(s)	井上, 沙織
Citation	北海道大学. 博士(歯学) 甲第11243号
Issue Date	2014-03-25
Doc URL	http://hdl.handle.net/2115/56121
Type	theses (doctoral - abstract of entire text)
Note	この博士論文全文の閲覧方法については、以下のサイトをご参照ください。
Note(URL)	https://www.lib.hokudai.ac.jp/dissertations/copy-guides/
File Information	Saori_inoue_summary.pdf



[Instructions for use](#)

学位論文内容の要約

学位論文題目

Multiwalled carbon nanotubes coating accelerates

osteoconductivity of anodized titanium

(多層カーボンナノチューブによる表面修飾は陽極酸化チ
タンの骨伝導性を促進する)

博士の専攻分野名称 博士（歯学） 氏名 井上 沙織

Multiwalled carbon nanotubes coating accelerates osteoconductivity of anodized titanium

Saori Inoue¹⁾, Tsukasa Akasaka²⁾, Fumio Watari²⁾, Atsuro Yokoyama¹⁾

- 1) Department of Oral Functional Prosthodontics Division of Oral Functional Science, Graduate School of Dental Medicine, Hokkaido University, Kita 13, Nishi 7, Kita-ku, Sapporo, Hokkaido, 060-8586, Japan. (Chief: Prof. Atsuro Yokoyama)
- 2) Department of Biomedical Materials and Engineering, Graduate School of Dental Medicine, Hokkaido University, Kita 13, Nishi 7, Kita-ku, Sapporo, Hokkaido, 060-8586, Japan.

Abstract

Because of their excellent mechanical, electrical and biocompatible properties, multiwalled carbon nanotubes (MWCNTs) are employed for tissue engineering. Recent studies have also revealed that CNTs provide a preferable surface for cell adhesion and growth. Osteoblast-like cells adhere well and grow on the MWCNT-coated collagen sponge better than on the non-coated sponge. On the other hand, titanium is commonly used as a biomaterial for dental implants because of its excellent mechanical and biocompatible properties. Various surface modifications, *e.g.* anodization and apatite coating, are employed for current dental implants to achieve earlier osseointegration. The purpose of this study was to coat anodized titanium with MWCNTs (CNT-Ti) and investigate the effects of this successfully CNT-Ti on osteoconductivity. Human osteosarcoma cells, Saos2, were cultured on CNT-Ti discs. CNT-Ti wires were implanted in the bone marrow space in femurs of rats for 2 and 4 weeks, and histological and histomorphometric evaluations were carried out. The anodized Ti could be coated homogeneously with MWCNTs. Saos2 cells on CNT-Ti showed prompted proliferation.

After 7 days, the DNA content of the cells on CNT-Ti was significantly higher than for other specimens. Animal experiments revealed that the bone contact ratio on the CNT-Ti was significantly higher than on the anodized Ti alone. Thus, surface modification with MWCNTs accelerated the osteoconductivity of anodized Ti.