



HOKKAIDO UNIVERSITY

Title	Impact of the fundamental use of surgical energy certification on surgeons' behavior and awareness of safe use of energy devices : a cross-sectional survey research
Author(s)	Kondo, Akihiro; Nishihara, Yuichi; Sato, Miho et al.
Citation	Surgical endoscopy and other interventional techniques, 37(1), 241-247 https://doi.org/10.1007/s00464-022-09468-4
Issue Date	2023-01-01
Doc URL	https://hdl.handle.net/2115/91063
Rights	This version of the article has been accepted for publication, after peer review (when applicable) and is subject to Springer Nature' s AM terms of use, but is not the Version of Record and does not reflect post-acceptance improvements, or any corrections. The Version of Record is available online at: http://dx.doi.org/10.1007/s00464-022-09468-4
Type	journal article
File Information	SEND-D-22-00143_R1.pdf



Impact of the Fundamental Use of Surgical Energy Certification on Surgeons'

Behavior and Awareness of Safe Use of Energy Devices: A Cross-Sectional Survey

Research

Running head: Surgical Safety Behavior and Awareness

Akihiro Kondo, MD, PhD¹, Yuichi Nishihara, MD, PhD², Miho Sato, BA³, Elif Bilgic,
PhD⁴, Yusuke Watanabe, MD, PhD^{3, 5, 6, *}

¹ Department of Gastroenterological Surgery, Faculty of Medicine, Kagawa University

² Department of Surgery, Shinmatsudo Central General Hospital

³ Clinical Research and Medical Innovation Center, Institute of Health Science

Innovation for Medical Care, Hokkaido University Hospital

⁴ Department of Experimental Surgery, McGill University

⁵ Center for Medical Device Development, Institute of Health Science Innovation for

Medical Care, Hokkaido University Hospital

⁶ Department of Gastroenterological Surgery II, Hokkaido University Faculty of

Medicine

***Correspondence to:** Yusuke Watanabe, MD, PhD

Department of Gastroenterological Surgery II

Hokkaido University Faculty of Medicine

Kita 14 Nishi 7, Kita-ku, Sapporo 060-8638, Japan

Tel: +81-11-706-7931

E-mail: yusuke.watanabe@pop.med.hokudai.ac.jp

Funding: The authors report no funding for this study.

Abstract

Background: The Fundamental Use of Surgical Energy (FUSE) program was established to educate surgeons and trainees to promote awareness and behaviors for the safe use of surgical energy devices. Despite its implementation, the impact of FUSE certification on surgeons' behavior and safety awareness regarding practice of energy devices remains unclear. This study aimed to identify the perceived impact of FUSE certification on surgeons' behavior and awareness regarding the safe use of surgical energy devices.

Methods: We performed a descriptive cross-sectional survey study, using non probabilistic purposive sampling, and distributed 22-item web-based questionnaires among all 59 FUSE-certified surgeons in Japan, excluding operating room nurses and medical students. The questionnaire items covered demographics, surgical techniques using various energy devices, changes in behavior and safety awareness, communication with colleagues about surgical energy devices, and educational activities related to energy devices.

Results: Fifty-seven participants completed the questionnaire (response rate 96.6%). Most surgeons (91.3%) could apply material learned from the FUSE program in practice, especially material related to monopolar electrosurgery. Fifty-six surgeons

(98.3%) reported increased awareness of surgical safety, and 35 (61.5%) reported increased communication with operating room personnel about the safe use of energy devices. Moreover, 56 participants (98.3%) indicated a need for systematic education in surgical energy, with participants recommending fellows (94.7% of participants specified that fellows should participate in further education), residents (75.4%), and attending surgeons (63.2%) as the target recipients of this training.

Conclusions: After FUSE certification, not only did surgeons' knowledge increase, but their energy-related surgical techniques in practice also improved. Furthermore, FUSE-certified surgeons felt that they were more aware of surgical-energy safety and were dedicated to its promotion.

Key Words: Fundamental Use of Surgical Energy (FUSE), certification, energy device, safe usage, surgeons' behavior, safety awareness

Introduction

Electrosurgery-related adverse events such as surgical burns, operating room (OR) fires, and unrecognized organ injuries can result in patient harm [1,2]. Therefore, surgeons should be well informed about the safe use of energy devices (e.g., monopolar electrosurgery, ultrasonically activated devices, or vessel sealing devices). However, recent data have shown that many surgeons, including those who are experienced, demonstrate knowledge gaps in the safe and effective use of energy devices [3,4]. To address this issue, the Society of American Gastrointestinal and Endoscopic Surgeons (SAGES) has established a web-based didactic curriculum, known as the Fundamental Use of Surgical Energy™ (FUSE) program [5]. In 2014, the FUSE program was introduced in Japan. The first hands-on seminar was held as a pre-congress workshop at the 27th annual meeting of the Japan Society for Endoscopic Surgery [6].

Previous studies have shown that a structured curriculum based on the FUSE program significantly increases surgical trainees' knowledge and self-perceived comfort with the safe use of electrosurgical devices, with retention after 3 months and 1 year [7,8]. Nonetheless, the impact of obtaining FUSE certification on surgeons' behavior in the OR and on changes in surgeons' safety awareness have not been investigated. Moreover, the extent to which institutions or communities have been influenced by

1
2
3 FUSE-certified surgeons in practice is unknown. Hence, the purpose of this study was
4
5
6 to explore the perceived impact of the FUSE program on surgeons' behavior and safety
7
8
9 awareness regarding the use of surgical energy devices.
10
11
12
13
14
15

16 **Materials and Methods**

17 18 19 Study Design and Population

20
21
22 We sent email invitations to each FUSE-certified surgeon in Japan (N = 59) through the
23
24
25 Japan Association for Surgical Education to participate in a cross-sectional survey in
26
27
28 March 2021. However, to adhere to the aim of investigating surgeons' behavior and
29
30
31 awareness, the OR nurses and medical students were excluded from this study.
32
33
34

35 Participants were asked to complete a web-based questionnaire developed using Google
36
37
38 Forms and were given 3 weeks to complete the questionnaire (responses were
39
40
41 anonymous). This study adhered to the Ethical Guidelines for Medical and Health
42
43
44 Research Involving Human Subjects in Japan [9]. Ethical approval by the institutional
45
46
47 review board of Kagawa University, Kagawa, Japan was exempted because this study
48
49
50 did not involve any patients, affect the health of participants, or use their personal
51
52
53 identification information. All participants provided written informed consent prior to
54
55
56 completing the survey.
57
58
59
60
61
62
63
64
65

Survey Design

The questionnaire had 22 items and was divided into the following sections:

demographics (4 items), surgeons' use of surgical techniques with various energy devices (4 items rated on 5-point Likert scale including a free description item), changes in surgeons' behavior and safety awareness (4 items rated on 5-point Likert scale including a free description item), frequency of communication with colleagues about electrosurgery-related topics (4 items rated on 5-point Likert scale), and educational activities related to energy devices that the surgeons participated in or conducted (6 items rated using single-choice, multiple choice, and 5-point Likert scale, including a free description item; Table 1). The questionnaire was developed by FUSE certified surgeons (AK, YN), who were experienced in designing educational activities on surgical energy devices; they listed potential items by exploring the degree of application among FUSE-certified individuals. Finally, the questionnaire was drafted through a discussion with surveyor/educational researcher (YW)—a SAGES FUSE committee member. The questionnaire items were validated by beta testing them with two FUSE-certified individuals. For scoring purposes, Likert-responses 1, 2, 3 were considered negative, while Likert-responses 4 and 5 were considered positive. Free description items required self-reported responses for the following: 1) contents that are

1
2
3 lacking in the FUSE program for clinical practice application, 2) content-changes
4
5
6 required for teaching usage of surgical energy devices, and 3) barriers in teaching the
7
8
9 FUSE program.
10

11 12 Statistical Analysis

13
14
15
16 The survey results were analyzed using descriptive statistics. Responses from
17
18
19 gastrointestinal surgeons and from physicians with other specialties were compared
20
21
22 using a Pearson's chi-squared test.
23
24
25
26
27

28 29 Results

30
31
32 Of the 59 FUSE-certified physicians in Japan, 56 surgeons and 1 anesthesiologist
33
34
35 completed the questionnaire (response rate 96.6%), as shown in Table 2. Almost half of
36
37
38 the participants were general surgeons (n=28, 49.1%), followed by gynecologists (n=17,
39
40
41 29.8%). Twenty-four (45.3%) participants had obtained FUSE certification within the
42
43
44 past 1–3 years, and 18 (31.6%) had obtained certification more than 3 years ago. Most
45
46
47 participants completed the FUSE examination owing to their interest in energy devices
48
49
50 (n=49, 86%) or to understand the proper use of energy devices (n=43, 75.4%).
51
52
53

54 However, there were no significant differences across item responses among surgical
55
56
57 specialties.
58
59
60
61
62
63
64
65

Surgical techniques using energy devices (Table 3)

Most participants (n=52, 91.3%) applied knowledge acquired from the FUSE program in clinical practice. Similarly, most participants acknowledged that they changed the way they used monopolar electrosurgery devices (n=50, 87.8%) and reported improvements in surgical techniques (n=47, 82.5%). In contrast, a positive change in perception was acknowledged by only approximately 50% of participants using ultrasonic activating devices (USADs) or vessel sealing systems (VSSs). The majority of participants reported a need for more detailed explanations within the FUSE content related to USAD as well as further explanations for VSS, specific energy devices, and advanced features of electrosurgical units (e.g., soft coag mode).

Changes in behavior and safety awareness (Table 4)

Almost all respondents (98.3%) reported an increase in surgical safety awareness, and 35 participants (61.5%) attempted to promote the safe use of energy devices at their institutions. For 46 (80.7%) participants, methods of teaching colleagues about energy devices changed. Based on the 34 descriptive answers offered by those who completed the survey, half of the respondents felt that changes occurred primarily when they were teaching surgical trainees about the fundamental principles and appropriate use of electrosurgery, as detailed in Supplemental Text 1.

1
2
3 Communications with surgeons and OR nurses (Table 5)
4
5

6 In terms of dispersive electrodes, while 38.6% of the participants had interactive
7
8 communication with surgeons or surgical trainees, participants had more opportunities
9
10 to communicate with OR nurses (49.2%). Overall, more than 70% of respondents had
11
12 increased communication with surgeons or nurses about the importance of cleaning
13
14 burnt materials (e.g., eschar) from the active electrode. The frequency of
15
16 communication with other surgeons/trainees and OR nurses about a topic was similar
17
18 for “electrosurgery setting” (61.4% and 59.6%, respectively) and “placement of the
19
20 handpiece (active electrode) when not in use” (43.9% and 45.6%, respectively).
21
22
23
24
25
26
27
28
29
30

31 Educational activities about energy devices (Table 6)
32
33

34 Fifty-six participants (98.2%) reported the need for systematic education in surgical
35
36 energy, with participants targeting fellows (94.7% of participants specified that fellows
37
38 should participate in further education), residents (75.4%), and attending surgeons
39
40 (63.2%) as recipients of this training. Thirty-one participants (54.4%) had some
41
42 experience teaching about energy devices to colleagues and OR staff, including didactic
43
44 lectures or workshops. Of the descriptive responses offered by 26 participants, 12
45
46 (46.2%) felt that senior surgeons’ lack of interest in the FUSE content created a barrier
47
48 to teaching activities.
49
50
51
52
53
54
55
56
57
58
59
60
61
62
63
64
65

Discussion

This study aimed to explore the impact of the FUSE program on surgeons' behavior and safety awareness regarding the use of surgical energy devices. Therefore, the current study survey responses suggested that participants perceived their behavior and awareness regarding safe use of energy devices to be improved after FUSE certification, regardless of surgical specialty. Most participants reported that post-certification, their promotion of safe use of energy devices to OR personnel increased and that their teaching approaches changed. Moreover, FUSE-certified surgeons have acted at various levels (institutional, regional, and national) to disseminate knowledge about the safe use of energy devices. This is the first study to explore the effect of FUSE certification on surgeons' behavioral changes and safety awareness in practice using Kirkpatrick's model, which is a widely used model to evaluate the results of educational curriculum [10].

There is a variety of evidence about the effectiveness of the FUSE program in fulfilling knowledge gaps about the safe use of energy devices [7,8,11]. Previous studies have shown that the knowledge and confidence levels of surgical trainees improved significantly after educational intervention based on FUSE; this knowledge and

1
2
3 confidence were retained after three months and after one year [7,8]. In other studies,
4
5
6 most trainees reported satisfaction with the program content and potential contributions
7
8
9 to patient safety, and that they would recommend the FUSE program to colleagues [11].
10
11
12 Although previous studies have reported the effectiveness of the FUSE program, the
13
14
15 degree of application and changes in individual clinical practice due to the FUSE
16
17
18 certification are unexplored. Therefore, this study aimed to bridge this gap by exploring
19
20
21 the individual changes in clinical practice for surgical safety after obtaining the FUSE
22
23
24 program certification.
25
26

27
28 FUSE-certified surgeons in this study consistently reported improvement in their
29
30
31 monopolar electrosurgery surgical techniques; some felt the need for more detailed
32
33
34 content regarding USADs. USAD content accounts for only approximately 6% of the
35
36
37 FUSE program, whereas monopolar electrosurgery content accounts for approximately
38
39
40 50%. While the FUSE program focuses on the fundamentals of the safe use of energy
41
42
43 devices, content regarding cavitation, drilling, and appropriate range of grasping the
44
45
46 tissue and its effect of thermal spread, which can potentially cause injury, are not
47
48
49 currently explained in detail. Since USADs are widely and more frequently used
50
51
52 devices in laparoscopic surgery, this information should be explained further in the
53
54
55 FUSE program.
56
57
58
59
60
61
62
63
64
65

1
2
3 The majority of FUSE-certified surgeons reported increased awareness of surgical
4
5
6 safety and more frequent communication with colleagues and OR nurses about
7
8
9 electrosurgery-related topics. These findings of our study may indicate surgeons'
10
11
12 improved safety behavior and awareness while using surgical energy devices in practice
13
14
15 after obtaining the FUSE certification. Accordingly, FUSE-certified surgeons could
16
17
18 increasingly contribute to enhancing the safety of patients who require surgical
19
20
21 treatment. However, the frequency of conversation among surgeons around dispersive
22
23
24 electrodes barely changed. This result may be because the majority of surgeons are not
25
26
27 likely to place a dispersive electrode by themselves. Nonetheless, appropriate placement
28
29
30 of the dispersive electrode should be taught to surgeons. While most respondents
31
32
33 reported experience, or intentions of, conducting educational activities, intra-operative
34
35
36 teaching activities of energy-device-related topics must be structured and optimized.
37
38
39
40

41 Importantly, almost half of the study participants felt that the lower interest of senior
42
43
44 colleagues in the FUSE content was a barrier to conducting teaching activities. The
45
46
47 majority of respondents also felt that both attending surgeons and surgical trainees
48
49
50 should receive training in FUSE. Attending surgeons must also learn the safe use of
51
52
53 energy devices, as previous studies have shown that surgeons have knowledge gaps in
54
55
56 the safe and effective use of energy devices, regardless of experience[3,4]. Therefore,
57
58
59
60
61
62
63
64
65

1
2
3 buy-in and support from senior surgeons regarding the FUSE program is crucial for
4
5
6 further dissemination of educational activities that promote a safety culture around
7
8
9 surgical energy. Multidisciplinary approaches involving anesthesiologists, OR nurses,
10
11
12 and clinical engineers are likely to play a significant role in improving the safe use of
13
14
15 energy devices.
16
17

18
19 A strength of this study is its excellent response rate, which yields a more accurate
20
21
22 interpretation of the results. However, the data should be interpreted with caution given
23
24
25 the potential of bias. This study only collected views of FUSE certified personnel, who
26
27
28 might be strongly motivated and interested in the FUSE certification. This study has
29
30
31 other limitations as well. First, although differences between surgical specialties might
32
33
34 have affected results, the differences were not adequately considered for interpretation.
35
36
37 Surgeons prefer to use a variety of energy devices, preferences that would change for
38
39
40 each specialty. However, as this study focused on more general topics of energy, tool
41
42
43 preference by specialty would not be a significant factor. Second, surgeons' training,
44
45
46 and clinical and teaching experiences vary; therefore, the impact of their experiences on
47
48
49 their responses must be accounted for when interpreting the results. Third, the current
50
51
52 study results may reflect a response bias because only self-reported data were collected
53
54
55 without any pre-tests before obtaining FUSE certification. Fourth, the generalizability
56
57
58
59
60
61
62
63
64
65

1
2
3 of this study may be limited, as only FUSE-certified surgeons in Japan were included.
4
5
6 Moreover, inferential statistics using pre- and post-FUSE certification tests would be
7
8
9 required to achieve more reliable results, and further survey studies are required with
10
11
12 international FUSE-certified surgeons. Fifth, the institutional impact of certified
13
14
15 surgeons was not investigated. Although certified surgeons reported that they changed
16
17
18 their practice, actual improvements in behavior and awareness regarding surgical energy
19
20
21 in the OR remain unclear. It also remains unclear if perceived changes occurred in other
22
23
24 areas. Finally, this was not a comparative study between the group of FUSE-certified
25
26
27 surgeons and the control group. FUSE is an educational program contributing to
28
29
30 improving OR safety. The rate of electrosurgery related adverse events needs to be
31
32
33 collected to demonstrate the definitive effectiveness of the educational program.
34
35
36 Although a comparative study is vital to present the definitive effectiveness of FUSE
37
38
39 certification, collecting the comparative data would be challenging, due to the immense
40
41
42 effort required and lack of feasibility.
43
44
45

46
47 FUSE-certified surgeons reported behavioral changes and increased awareness of
48
49
50 the safety of applying energy devices in their practice. They were more open to
51
52
53 conversation and teaching about the fundamentals of energy devices with and to other
54
55
56 OR personnel. Study findings suggest that FUSE certification has positive effects on the
57
58
59
60
61
62
63
64
65

1
2
3 safe use of energy devices in the OR. Based on these findings that FUSE certification
4
5
6 could achieve behavioral change and improved consciousness to use energy devices
7
8
9 safely and appropriately, our goal is to disseminate the FUSE program or its concept
10
11
12 among all surgeons to promote safe surgery.
13
14
15
16
17
18

19 **Acknowledgements**

20
21
22 The authors thank all FUSE-certified personnel in Japan and, especially, the participants
23
24
25 in this study. This study would not have been possible without the generosity of all
26
27
28 participants.
29
30
31
32
33
34

35 **Disclosures**

36
37
38 **Conflict of Interest:** Dr. Yusuke Watanabe received honoraria from Medtronic, Johnson
39
40
41 & Johnson, Olympus, ConMed, and AMCO outside the submitted study. Dr. Akihiro
42
43
44 Kondo, Dr. Yuichi Nishihara, Ms. Elif Bilgic, and Ms. Miho Sato have no conflicts of
45
46
47 interest or financial ties to disclose.
48
49
50
51
52
53
54
55
56
57
58
59
60
61
62
63
64
65

References

1. Aigner N, Fialka C, Fritz A, Wruhs O, Zöch G (1997) Complications in the use of diathermy. *Burns* 23(3):256-264. DOI: [10.1016/s0305-4179\(96\)00113-1](https://doi.org/10.1016/s0305-4179(96)00113-1).
2. Nduka CC, Super PA, Monson JR, Darzi AW (1994) Cause and prevention of electrosurgical injuries in laparoscopy. *J Am Coll Surg* 179(2):161-170.
3. Feldman LS, Fuchshuber P, Jones DB, Mischna J, Schwaitzberg SD, FUSE (2012) Surgeons don't know what they don't know about the safe use of energy in surgery. *Surg Endosc* 26(10):2735-2739. DOI: [10.1007/s00464-012-2263-y](https://doi.org/10.1007/s00464-012-2263-y).
4. Watanabe Y, Kurashima Y, Madani A, Feldman LS, Ishida M, Oshita A, Naitoh T, Noma K, Yasumasa K, Nagata H, Nakamura F, Ono K, Suzuki Y, Matsuhashi N, Shichinohe T, Hirano S (2016) Surgeons have knowledge gaps in the safe use of energy devices: a multicenter cross-sectional study. *Surg Endosc* 30(2):588-592. DOI: [10.1007/s00464-015-4243-5](https://doi.org/10.1007/s00464-015-4243-5).
5. Madani A, Jones DB, Fuchshuber P, Robinson TN, Feldman LS (2014) Fundamental Use of Surgical Energy™ (FUSE): a curriculum on surgical energy-based devices. *Surg Endosc* 28(9):2509-2512. DOI: [10.1007/s00464-014-3623-6](https://doi.org/10.1007/s00464-014-3623-6).
6. Fuchshuber P, Schwaitzberg S, Jones D, Jones SB, Feldman L, Munro M, Robinson T, Purcell-Jackson G, Mikami D, Madani A, Brunt M, Dunkin B, Gugliemi C,

Groah L, Lim R, Mischna J, Voyles CR (2018) The SAGES Fundamental Use of Surgical Energy program (FUSE): history, development, and purpose. Surg Endosc 32(6):2583-2602. DOI: [10.1007/s00464-017-5933-y](https://doi.org/10.1007/s00464-017-5933-y).

7. Madani A, Watanabe Y, Vassiliou MC, Fuchshuber P, Jones DB, Schwaitzberg SD, Fried GM, Feldman LC (2014) Impact of a hands-on component on learning in the Fundamental Use of Surgical Energy™ (FUSE) curriculum: a randomized-controlled trial in surgical trainees. Surg Endosc 28(10):2772-2782. DOI: [10.1007/s00464-014-3544-4](https://doi.org/10.1007/s00464-014-3544-4).

8. Madani A, Watanabe Y, Vassiliou MC, Fuchshuber P, Jones DB, Schwaitzberg SD, Fried GM, Feldman LS (2016) Long-term knowledge retention following simulation-based training for electrosurgical safety: 1-year follow-up of a randomized controlled trial. Surg Endosc 30(3):1156-1163. DOI: [10.1007/s00464-015-4320-9](https://doi.org/10.1007/s00464-015-4320-9).

9. Ministry of Health, Labour and Welfare. Ethical guidelines for medical and Health Research involving human subjects. Available at: <https://www.mhlw.go.jp/file/06-Seisakujouhou-10600000-Daijinkanboukouseikagakuka/0000080278.pdf>. Accessed 29 January 2021.

10. Kirkpatrick DL, Kirkpatrick JD (2006) Evaluating Training Program. The Four
Levels. 3rd ed. Berrett-Koehler Publishers, San Francisco.
11. Gugenheim J, Debs T, Gravié JF, Deleuze A, Millat B, Borie F, Mathonnet M (2020)
Results of the FUSE evaluation project in France. Surg Endosc 34(4):1819-1822.
DOI: [10.1007/s00464-019-06938-0](https://doi.org/10.1007/s00464-019-06938-0).

Table 1. Questionnaire content

Section	No. of items	Type of questions and scales
1. Participant demographics	4	-
2. Surgical techniques using various energy devices	4	Likert-scale: (1) Strongly disagree - (5) Strongly agree (1) Definitely not changed/improved - (5) Definitely changed/improved Free description
3. Changes in behavior and safety awareness	4	Likert-scale: (1) Definitely not changed - (5) Definitely changed Free description
4. Frequency of communications with colleagues about electrosurgery-related topics	4	Likert-scale: (1) Definitely not changed - (5) Every case
5. Educational activities on energy devices	6	Single-choice, Multiple-choice Likert-scale: (1) Strongly disagree - (5) Strongly agree Free description

Table 2. Participant demographics (n = 57)

	n (%)
Specialties	
General surgery	28 (49)
Gynecology	17 (29.8)
Urology	4 (7)
Pediatric surgery	4 (7)
Thoracic surgery	1 (1.8)
Endocrine surgery	1 (1.8)
Otolaryngology	1 (1.8)
Anesthesiology	1 (1.8)
Reasons for learning about energy devices via the FUSE program ^a	
Professional curiosity about energy devices	49 (86)
Learning the proper use of energy devices	43 (75.4)
Teaching surgical trainees	19 (33.3)
Years since date of FUSE certification	
< 0.5 year	8 (14)
0.5–1 year	7 (12.2)
1–3 years	24 (42.1)
> 3 years	18 (31.6)

Table 3. Self-reported changes in surgical technique using each energy device (n = 57)

	Strongly disagree				Strongly agree
	1	2	3	4	5
	n (%)				
Changes in the use of energy devices in practice	0 (0)	1 (1.8)	4 (7)	25 (43.9)	27 (47.4)
	Definitely not changed				Definitely changed
	1	2	3	4	5
	n (%)				
Changes in surgical technique:					
Monopolar electrosurgery	0 (0)	1 (1.8)	6 (10.5)	12 (21.1)	38 (66.7)
USAD	3 (5.3)	4 (7)	19 (33.3)	25 (43.9)	6 (10.5)
VSS	2 (3.5)	4 (7)	21 (36.8)	23 (40.4)	7 (12.3)
	Definitely not improved				Definitely improved
	1	2	3	4	5
	n (%)				
Improvements in surgical technique:					
Monopolar electrosurgery	0 (0)	0 (0)	10 (17.5)	23 (40.4)	24 (42.1)
USAD	2 (3.5)	5 (8.8)	20 (35.1)	28 (49.1)	2 (3.5)
VSS	2 (3.5)	4 (7)	20 (35.1)	28 (49.1)	3 (5.3)

USAD = ultrasonic activating device; VSS = vessel sealing system

Table 4. Self-reported changes in behavior and safety awareness (n = 57)

	Definitely not changed				Definitely changed
	1	2	3	4	5
	n (%)				
Awareness of surgical safety	0 (0)	0 (0)	1 (1.8)	25 (43.9)	31 (54.4)
Content of the teaching	0 (0)	0 (0)	11 (19.3)	26 (45.6)	20 (35.1)
	Definitely not approached				Definitely approached
	1	2	3	4	5
	n (%)				
Approaches to promote the safe use of energy devices in each institution	1 (1.8)	4 (7)	17 (29.8)	23 (40.4)	12 (21.1)

Table 5. Frequency of communication with surgeons and operating room nurses about the following topics (n=57)

	With surgeons / ORNs	Definitely not changed				Every case
		1	2	3	4	
n (%)						
Electrosurgery setting	Surgeons	4 (7)	6 (10.5)	12 (21.1)	26 (45.6)	9 (15.8)
	ORNs	4 (7)	7 (12.3)	12 (21.1)	28 (49.1)	6 (10.5)
Dispersive electrode	Surgeons	6 (10.5)	9 (15.8)	20 (35.1)	17 (29.8)	5 (8.8)
	ORNs	7 (12.3)	6 (10.5)	16 (28.1)	23 (40.4)	5 (8.8)
Placement of the handpiece (active electrode) when not in use	Surgeons	5 (8.8)	12 (21.1)	15 (26.3)	20 (35.1)	5 (8.8)
	ORNs	8 (14)	7 (12.3)	16 (28.1)	24 (42.1)	2 (3.5)
Cleaning burnt materials (e.g., eschar) from active electrode	Surgeons	4 (7)	5 (8.8)	8 (14)	25 (43.9)	15 (26.3)
	ORNs	3 (5.3)	6 (10.5)	7 (12.3)	27 (47.4)	14 (24.6)

ORN = operating room nurse
Surgeons include senior or junior surgeons and surgical trainees

Table 6. Energy device educational activities (n=57)

	n (%)
Participated in FUSE workshop	
Yes	45 (78.9)
No	12 (21.1)
Need for systematic education in surgical energy	
5 strongly agree	46 (80.7)
4	10 (17.5)
3	1 (1.8)
2	0 (0)
1 strongly disagree	0 (0)
Preferred targets for education ^a	
Medical students	21 (35.8)
Residents	43 (75.4)
Fellows	54 (94.7)
Attending surgeons	40 (70.2)
Attending surgeons with teaching responsibilities	36 (63.2)
Experience teaching surgical energy (e.g., didactic lecture or workshop)	
Yes	31 (54.4)
No	26 (45.6)
If “No,” intention of conducting teaching activities	
Yes	21 (80.7)
No intention	1 (3.9)
Unknown	4 (15.4)

FUSE = Fundamental Use of Surgical Energy

^aRespondents could choose more than one response

Supplementary Material

Supplemental Text 1. Self-reported content changes teaching about surgical-energy devices (34 descriptive responses summarized)

- Logical explanations based on principles
- Teaching with hands-on demonstrations
- Explanations of practical use focusing on tissue effect, activation time
- Importance of lower voltage output
- Use low voltage continuous outputs (“cut mode output”) to seal vessels
- Applications and practical use of the cut mode output
- Differences in output modes (eg, cut mode vs. coag mode)/power and practical applications of each mode
- Lateral thermal spread among output modes
- Positions of dispersive electrodes
- Potential adverse effect of electrosurgery (eg, residual heat injury, capacitive coupling, operating room fire) and how to avoid it
- Avoid open activation (not activating the electrode before it becomes in contact with the targeted tissue)
- Controlling current density by changing the contact area of active electrode