



HOKKAIDO UNIVERSITY

Title	Tele-robotic distal gastrectomy with lymph node dissection on a cadaver
Author(s)	Ebihara, Yuma; Hirano, Satoshi; Kurashima, Yo et al.
Citation	Asian journal of endoscopic surgery, 17(1), e13246 https://doi.org/10.1111/ases.13246
Issue Date	2023-09-19
Doc URL	https://hdl.handle.net/2115/94183
Rights	This is the peer reviewed version of the following article: Ebihara Y, Hirano S, Kurashima Y, et al. Tele-robotic distal gastrectomy with lymph node dissection on a cadaver. Asian J Endosc Surg. 2024; 17(1):e13246., which has been published in final form at https://doi.org/10.1111/ases.13246 . This article may be used for non-commercial purposes in accordance with Wiley Terms and Conditions for Use of Self-Archived Versions. This article may not be enhanced, enriched or otherwise transformed into a derivative work, without express permission from Wiley or by statutory rights under applicable legislation. Copyright notices must not be removed, obscured or modified. The article must be linked to Wiley's version of record on Wiley Online Library and any embedding, framing or otherwise making available the article or pages thereof by third parties from platforms, services and websites other than Wiley Online Library must be prohibited.
Type	journal article
File Information	AJES_ases.13246.pdf



Short Report

Title: Tele-robotic distal gastrectomy with lymph node dissection on a cadaver

Authors

Yuma Ebihara¹⁾²⁾ (ORCID; 0000-0002-7317-8067), Satoshi Hirano¹⁾²⁾ (co-first author), Yo Kurashima²⁾,

Hironobu Takano²⁾, Kunishige Okamura²⁾, Soichi Murakami²⁾³⁾, Toshiaki Shichinohe²⁾³⁾, Hajime

Morohashi¹⁾⁴⁾, Eiji Oki¹⁾⁵⁾, Kenichi Hakamada¹⁾⁴⁾, Norihiko Ikeda¹⁾⁶⁾, Masaki Mori¹⁾⁷⁾

Affiliations

1) Committee for Promotion of Remote Surgery Implementation, Japan Surgical Society, Tokyo, Japan

2) Department of Gastroenterological Surgery II, Hokkaido University Faculty of Medicine, Sapporo,

Japan

3) Center for Education Research and Innovation of Advanced Medical Technology, Hokkaido University

Hospital, Sapporo, Japan

4) Department of Gastroenterological Surgery, Hirosaki University Graduate School of Medicine,

Hirosaki, Japan

5) Department of Surgery and Science, Kyushu University, Fukuoka, Japan

6) Department of Surgery, Tokyo Medical University, Tokyo, Japan

7) Tokai University School of Medicine, Isehara, Japan

Corresponding author

Satoshi Hirano, MD, PhD

Department of Gastroenterological Surgery II, Hokkaido University Faculty of Medicine.

North 15 West 7, Kita-ku, Sapporo 0608638, Hokkaido, Japan

Tel: +81 11 706 7714 Fax: +81 11 706 7158

Email: satto@med.hokudai.ac.jp

ABSTRACT

The purpose of this study is to evaluate the performance of tele-robotic distal gastrectomy (tele-RDG) with lymph node dissection (LND) using a novel Japanese-made surgical robot hinotori™ (Medicaroid, Kobe, Japan) in a cadaver with a presumptive gastric cancer. The Cadaveric Anatomy and Surgical Training laboratory (CAST-Lab.) at Hokkaido University and Kushiro City General Hospital (KCGH) are connected by a guaranteed type line (1 Gbps), and the distance between the two facilities is 250 km. A patient cart was installed at CAST-Lab, and a surgeon cockpit was installed at KCGH. Tele-RDG with D2 LND was performed on an adult human cadaver. In all surgical processes, the communication environment was stable without image degradation, and the mean RTT was 40 ms (36.5-55 ms). For tele-RDG with D2 LND, the operation time was 199 min. without any technical problems. Tele-RDG using hinotori™ was feasible and similar to local robotic RDG.

Keywords: Telesurgery, robotic surgery, gastrectomy, cadaver

Introduction

In recent years, the development of high-speed, high-capacity communication technologies using optical fiber and 5th generation mobile communication systems (5G), and new surgical robots have made remote surgery a reality¹. The Japan Surgical Society (JSS) has published a series of studies to establish the minimum requirements for telesurgery, especially the telesurgical support needed to perform actual telesurgery using a novel Japanese-made surgical robot hinotoriTM (Medicaroid, Kobe, Japan)^{2,3}. The purpose of this study is to evaluate the surgical feasibility of tele-robotic distal gastrectomy (tele-RDG) with lymph node dissection (LND) using hinotoriTM on a cadaver with presumptive gastric cancer in a clinical setting.

Material and Methods

The Cadaveric Anatomy and Surgical Training laboratory (CAST-Lab.) at Hokkaido University and Kushiro City General Hospital (KCGH), which are 250 km apart, were connected by a guaranteed type line (1G-bps) provided by NTT East (Nippon Telegraph and Telephone East Corporation, Tokyo, Japan) using Internet Protocol Security (IPSec) encryption. The operating unit was installed in the CAST-Lab. A surgeon cockpit was installed at KCGH, where the surgeon performed the tele-RDG with LND. Time (RTT; in milliseconds) where the round trip of communication was measured (Fig.1). The request was sent by a switch on the surgeon's cockpit side, and the response time from the switch on the operator's side was measured intraoperatively. The study was conducted using a Thiel-embalmed cadaver (male)

with an age at death of 80 years. The surgery was conducted by a physician certified in the robot proctoring qualification system of the Japan Society for Endoscopic Surgery (JSES), and the evaluators were skilled surgeons who were JSES technical certified physicians. We performed tele-RDG with D2 LND and intracorporeal Billroth I reconstruction in the cadaver (Fig. 2). The regions of lymph node dissection were defined according to the Japanese Classification of Gastric Carcinoma⁴. Details of this procedure have been described in our previous report on human patients⁵. Two assisting surgeons and one scrub nurse were assigned to CAST-Lab and one surgeon was assigned to KCGH to perform tele-RDG. We measured the operation time, Global Evaluative Assessment of Robotic Skills (GEARS)⁶, Global Operative Assessment of Laparoscopic Skills (GOALS)⁷, Robot Usability Score (RUS)⁸, system Usability Scale (SUS)⁹, and Piper Fatigue Scale-12 (PFS-12)¹⁰. GEARS and GOALS consist of each domain (depth perception, bimanual dexterity, efficiency, force sensitivity, autonomy, and robot control) and are scored on a 5-point scale to reliably quantify technical proficiency. The technical evaluation was conducted by two physicians certified in the endoscopic surgical skill qualification system of JSES¹¹.

Results

In all surgical processes, the communication environment was stable without image degradation, and the mean RTT was 40 ms. (36.5-55 ms). Tele-RDG with D2 LND was performed safely without organ injury. The operation time was 199 minutes, The overall score were GEARS:28 (depth perception:5,

bimanual dexterity:5, efficiency:4, force sensitivity 5, autonomy:5, robotic control:4), GOALS:23 (depth perception:5, bimanual dexterity:5, efficiency:4, tissue handling:4, autonomy:5), RUS:34 (physical comfort:4, hand controls:4, foot controls:4, 3D vision:5, annoyed or stressed:4, smooth:4, satisfaction:4, actuality:5), SUS:82.5 (acceptability ranges: acceptable, grade scale: A, adjective ratings: excellent), and PFS-12:26 (behavioral:9, affective:6, sensory:6, cognitive:5).

Discussion

Our study revealed that tele-RDG with LND using the hinotori™ surgical robot system is feasible for local robotic surgery. To the best of our knowledge, this is the first report of a tele-RDG with LND envisioned for gastric cancer in a cadaver.

In 2001, Marescaux et al. reported the first successful completion of transatlantic robot-assisted cholecystectomy using a Zeus surgical robot (formerly Computer Motion, CA) in clinical practice¹². One advantage of tele-robotic surgery (tele-RS) is that it reduces physical, emotional, and financial burdens by eliminating the need for patients and surgeons to travel. In areas where medical resources are scarce, the potential to contribute to the population through tele-RS is significant. In comparison to open or conventional laparoscopic surgery, RS having the medical staff away from the bedside potentially reduces not only contamination with body fluids and surgical gasses of the surgical area but also the number of directly exposed medical staff¹³. Tele-RS can play a pivotal role in surgical procedures performed during

the pandemic caused by the currently prevalent coronavirus disease.

In this study, the communication environment was stable, with an average RTT of 40 ms on a guaranteed line (1 Gbps), and all surgical procedures, including gastrectomy, lymph node dissection, and intracorporeal anastomosis, were performed without any problems caused by communication delay. Regarding transmission delay, we reported that operability decreases when the delay time perceived by the surgeon exceeds 100 ms in dynamic environments¹⁴, and on hemostasis and cholecystectomy in animal experiments. In the future, we plan to validate delays in more sensitive tasks (e.g. lymphadenectomy and reconstruction) in dynamic environments.

Regarding the tele-RDG with D2 LND, the operation time was 199 min. and was performed without any organ injury nor technical trouble. Moreover, the technical evaluation score (GEARS:28, GOALS:23), usability score (RUS:34, SUS:82.5), and surgeon fatigue score (PFS-12:26) were similar to those of clinical RDG.

This study had several limitations. The current experiment was conducted by only one expert surgeon, and its introduction into clinical practice should be made with caution. Clinical trials are necessary before tele-RDG with LND can be introduced into clinical practice in the future.

Conclusion

Tele-RDG with LND using hinotori™ on cadavers was feasible and is similar to the local robotic

RDG for gastric cancer.

Acknowledgements and Disclosure:

Acknowledgments

This study was conducted as part of the Advanced Telemedicine Network Research Project of the Japan Agency for Medical Research and Development (AMED), for which we are grateful for the financial support. We thank all staff of the Cadaveric Anatomy and Surgical Training Laboratory (CAST Lab). and engineers participating in this study. We also express our deepest gratitude to the Nippon Telegraph And Telephone East Corporation and Medicaroid for their cooperation during the experiments. We would like to thank Editage (www.editage.com) for the English language editing.

Funding: This work was supported by a grant from the Japan Agency for Medical Research and Development (AMED) (Grant Number 21hs0122001h0002).

Competing interests: The authors have no relevant financial or nonfinancial interests to disclose.

Toshiaki Shichinohe is an Editorial Board member of ASES Journal and a co-author of this article. To minimize bias, they were excluded from all editorial decision-making related to the acceptance of this article for publication.

Author contributions: All authors contributed to the conception and design of the study. Yuma Ebihara, Hironobu Takano, and Satoshi Hirano helped in the material preparation, data collection, and analysis.

The first draft of the manuscript was written by Yuma Ebihara and Satoshi Hirano, and all authors commented on the previous versions of the manuscript. All the authors have read and approved the final manuscript.

Ethics approval and consent to participate: Ethical Review Board for Life Science and Medical Research, Hokkaido University Hospital approved the cadaveric study, data collection and analysis (No. 022-0363). All managements of the cadaver study including the informed consent of body donation and opt-out of this research and were carried out in accordance with guidelines for cadaver dissection in education and research of clinical medicine by Japan Surgical Society and Japanese Association of Anatomists.

Consent to publish: All authors confirmed that consent to publish has been obtained.

References

1. Hakamada K, Mori M. The changing surgical scene: from the days of Billroth to the upcoming future of artificial intelligence and telerobotic surgery. *Ann Gastroenterol Surg* (2021);5:268–269.
2. Ebihara Y, Oki E, Hirano S, Takano H, Ota M, Morohashi H, Hakamada K, Urushidani S, Mori M. Tele-assessment of bandwidth limitation for remote robotics surgery. *Surg Today* (2022);12:1–7.
3. Takahashi Y, Hakamada K, Morohashi H, Akasaka H, Ebihara Y, Oki E, Hirano S, Mori M. Reappraisal of telesurgery in the era of high-speed, high-bandwidth, secure communications: evaluation

of surgical performance in local and remote environments. *Ann Gastroenterol Surg* (2023);7:167–174.

4. Japanese Gastric Cancer Association. Japanese gastric cancer treatment guidelines 2014 (ver.

4). *Gastric Cancer* (2017) 20:1–19.

5. Ebihara Y, Kurashima Y, Murakami S, Shichinohe T, Hirano S. Short-term outcomes of robotic distal gastrectomy with the “preemptive retropancreatic approach”: a propensity score matching analysis.

J Robot Surg (2022);16:825–831.

6. Aghazadeh MA, Jayaratna IS, Hung AJ, Pan MM, Desai MM, Gill IS, Goh AC. External validation of global evaluative assessment of robotic skills (GEARS). *Surg Endosc* (2015);29:3261–3266.

7. Vassiliou MC, Feldman LS, Andrew CG, Bergman S, Leffondré K, Stanbridge D, Fried GM. A global assessment tool for evaluation of intraoperative laparoscopic skills. *Am J Surg* (2005);190:107–113.

8. Tanaka A, Graddy C, Simpson K, Perez M, Truong M, Smith R. Robotic surgery simulation validity and usability comparative analysis. *Surg Endosc* (2016);30:3720–3729.

9. Bangor PT. Determining What Individual SUS Scores Mean: Adding an Adjective Rating Scale Kortum, and J.T. Miller. *J Usability Stud* (2009);4:114–123

10. Reeve BB, Stover AM, Alfano CM, Smith AW, Ballard-Barbash R, Bernstein L, McTiernan A, Baumgartner KB, Piper BF. The piper fatigue scale-12 (PFS-12): psychometric findings and item reduction in a cohort of breast cancer survivors. *Breast Cancer Res Treat* (2012);136:9–20.

11. Japanese Society for Endoscopic Surgery. Guidelines for the management of endoscopic surgery. Tokyo: *Japanese Soc for Endosc Surg*; 2014.
12. Marescaux J, Leroy J, Gagner M, Rubino F, Mutter D, Vix M, Butner SE, Smith MK. Transatlantic robot-assisted telesurgery. *Nature* (2001);413:379–380.
13. Kimmig R, Verheijen RHM, Rudnicki M; for SERGS Council. Robot assisted surgery during the COVID-19 pandemic, especially for gynecological cancer: a statement of the Society of European Robotic Gynaecological Surgery (SERGS). *J Gynecol Oncol*. 2020;31:e59. doi: 10.3802/jgo.2020.31.e59. Epub 2020 Apr 3. PMID: 32242340; PMCID: PMC7189073.
14. Nankaku A, Tokunaga M, Yonezawa H, et al. Maximum acceptable communication delay for the realization of telesurgery. *PLOS ONE*. 2022;17:e0274328. <https://doi.org/10.1371/journal.pone.0274328>, PMID: 36201429, PMCID: PMC9536636

Figure legends:

Fig.1


(Up) Operation unit (OU) at the clinical autopsy training center (CAST-Lab.) at Hokkaido University (HU) and surgeon cockpit (SC) in the remote site 250 km away (KCGH; Kushiro City General Hospital) where the surgeon performed tele-robotic distal gastrectomy with D2 lymph node dissection using a novel Japanese-made surgical robot hinotori™ (Medicaroid, Kobe, Japan) on a cadaver. (Down) OU at CAST-Lab and SC at KCGH were connected by a guaranteed type line (G-1Gbps) provided by NTT East (Nippon Telegraph and Telephone East Corporation, Tokyo, Japan) using Internet Protocol Security (IPSec) encryption. Round trip time (RTT) is composed of communication line delay.

Fig.2

A: Operation unit at the clinical autopsy training center at Hokkaido University. B: surgeon cockpit at Kushiro City General Hospital. C: Final view of the dissection of suprapancreatic lymph nodes. (CHA, common hepatic artery; GDA, gastroduodenal artery; LGA, left gastric artery; PHA, proper hepatic artery; SA, splenic artery; RGA, right gastric artery) D: Final view of the intracorporeal Billroth-I

reconstruction. (Duo, duodenum; St. stomach)

Operation unit (OU)



CAST-Lab. at HU

G-1Gbps
(250km)

KCGH





Surgeon cockpit (SC)

Guaranteed type line (1-Gbps)



Operation unit



Surgeon cockpit

